(19) Europäisches Patentamt
European Patent Office
Office européen des brevets

(11) EP 1 169 975 A1

(12) EUROPEAN PATENT APPLICATION

(43) Date of publication:
09.01.2002 Bulletin 2002/02

(51) Int Cl.[7]: **A61B 18/14**

(21) Application number: **01305874.8**

(22) Date of filing: **06.07.2001**

(84) Designated Contracting States:
AT BE CH CY DE DK ES FI FR GB GR IE IT LI LU MC NL PT SE TR
Designated Extension States:
AL LT LV MK RO SI

(30) Priority: **07.07.2000 US 611849**

(71) Applicant: **Biosense Webster, Inc.**
Diamond Bar, California 91765 (US)

(72) Inventors:
- **Mest, Robert A.**
Long Beach, California 90808 (US)
- **Nakagawa, Hiroshi**
Edmond, Oklahoma 73003 (US)

(74) Representative: **Mercer, Christopher Paul et al**
Carpmaels & Ransford 43, Bloomsbury Square
London WC1A 2RA (GB)

(54) Catheter with tip electrode having a recessed ring electrode mounted thereon

(57) A catheter particularly suitable for bipolar mapping and ablating comprises an elongated flexible body having a distal region and at least one lumen extending therethrough. A tip electrode having an outer diameter is mounted on the distal region. A proximal recessed central region has a diameter less than the tip. A ring electrode, isolated from the tip is mounted on the recessed central region. The ring electrode has an outer diameter less than the outer diameters of the tip. With this design, the exposed region of the tip electrode is in direct contact with the heart tissue, and thus senses both the local activation energy (near-field signals) at the point of contact with the heart tissue and far field activation energy (far-field signals) received by the exposed region through the blood. However, the recessed electrode is protected from direct contact with the heart tissue, but does contact with surrounding blood. The close proximity of the recessed electrode to the exposed region enables the recessed electrode to receive approximately the same far-field signals as the exposed region. However, the recessed electrode does not pick up the local activation potential (near-field signals) that are received by the exposed region. This design permits the creation of high resolution electrograms.

EP 1 169 975 A1

Description

FIELD OF THE INVENTION

[0001] The present invention is directed to a catheter for diagnosis and treatment of the heart, and more particularly to a catheter comprising a tip electrode having a recessed ring electrode mounted thereon.

BACKGROUND OF THE INVENTION

[0002] Electrode catheters have been in common use in medical practice for many years. They are used to stimulate and map electrical activity in the heart and to ablate sites of aberrant electrical activity.

[0003] In use, the electrode catheter is inserted into a major vein or artery, e.g., femoral artery, and then guided into the chamber of the heart which is of concern. Once the catheter is positioned within the heart, the location of aberrant electrical activity within the heart is then located.

[0004] One location technique involves an electrophysiological mapping procedure whereby the electrical signals emanating from the conductive endocardial tissues are systematically monitored and a map is created of those signals. By analyzing that map, the physician can identify the interfering electrical pathway. A conventional method for mapping the electrical signals from conductive heart tissue is to percutaneously introduce an electrophysiology catheter (electrode catheter) having mapping electrodes mounted on its distal extremity. The catheter is maneuvered to place these electrodes in contact with or in close proximity to the endocardium. By monitoring the electrical signals at the endocardium, aberrant conductive tissue sites responsible for the arrhythmia can be pinpointed.

[0005] For mapping, it is desirable to have a relatively small mapping electrode. It has been found that smaller electrodes record more accurate and discrete electrograms. Additionally, if a bipolar mapping arrangement is used, it is desirable that the two electrodes of the mapping arrangement be in close proximity to each other and that they be similar in size to produce more accurate and useful electrograms.

[0006] Additionally, it has been found that standard bipolar mapping catheters have limited accuracy because both electrodes pick up near field electrical signals emanating from the conductive endocardial tissues due to their contact with the heart tissue, and far-field electrical signals that propagate from other regions of the heart due to their contact with the blood. The far-field signals interfere with the near-field signals, making accurate measurement of the near-field signals difficult.

[0007] Once the origination point for the arrhythmia has been located in the tissue, the physician uses an ablation procedure to destroy the tissue causing the arrhythmia in an attempt to remove the electrical signal irregularities and restore normal heart beat or at least an improved heart beat. Successful ablation of the conductive tissue at the arrhythmia initiation site usually terminates the arrhythmia or at least moderates the heart rhythm to acceptable levels.

[0008] A typical ablation procedure involves providing a reference electrode, generally taped to the skin of the patient. RF (radio frequency) current is applied to the tip electrode, and current flows through the media that surrounds it, i.e., blood and tissue, toward the reference electrode. The distribution of current depends on the amount of electrode surface in contact with the tissue as compared to blood, which has a higher conductivity than the tissue. Heating of the tissue occurs due to its electrical resistance. The tissue is heated sufficiently to cause cellular destruction in the cardiac tissue resulting in formation of a lesion within the cardiac tissue which is electrically non-conductive. During this process, heating of the electrode also occurs as a result of conduction from the heated tissue to the electrode itself. If the electrode temperature becomes sufficiently high, a thin transparent coating of dehydrated blood protein can form on the surface of the electrode. If the temperature continues to rise, this dehydrated layer can become progressively thicker, resulting in char and/or thrombus on the electrode surface. The creation of char and thrombus is unsafe, as the char and thrombus can be dislodged from the electrode during the procedure or during removal of the catheter after the procedure.

[0009] In clinical practice, it is desirable to reduce or eliminate the formation of char and thrombus and, for certain cardiac arrhythmias, to create larger and/or deeper lesions. One method for accomplishing this end is to monitor the temperature of the ablation electrode and to control the RF current delivered to the ablation electrode based on this temperature. If the temperature rises above a preselected value, the current is reduced until the temperature drops below this value.

[0010] Another method for determining whether char and thrombus is forming is by monitoring the impedance. Specifically, because dehydrated biological material has a higher electrical resistance than endocardial tissue, impedance to the flow of electrical energy into the tissue also increases. A significant impedance rise thus indicates the formation of char and/or thrombus. With a relatively large electrode, e.g., an 8 mm electrode, an impedance rise from char and thrombus formation may not be easily detected because the char and thrombus is formed over a relatively small percentage of the total surface area of the electrode. In contrast, if the electrically active surface area of the electrode is relatively small, char and thrombus will form over a relatively larger area of the electrode, making detection by impedance measurements easier.

[0011] The cooling effect of the blood on the electrode is dependent on the thermal properties of the ablation electrode, and in particular its surface area. Typically ablation electrodes are relatively long, most commonly at least about 4 mm and up to about 8 mm, to provide suf-

ficient surface area for cooling. However, such electrodes are less suitable for mapping because, as discussed above, more accurate electrograms can be obtained with a smaller mapping electrode. It is desirable for electrophysiologists use the same catheter for mapping and ablation during a single procedure because, once a site is identified with a high resolution mapping electrode catheter as a target for therapy, it would be difficult to locate that site again with another catheter, particularly where the other catheter has a larger tip electrode. Accordingly, a need exists for an electrode catheter that has a relatively small surface area for enhanced mapping, but has good thermal properties for enhanced cooling during ablation.

[0012] Another disadvantage of typical ablation electrodes is that it is sometimes difficult to accurately predict lesion size because the lesion size can vary depending on the orientation of the ablation electrode. For example, typically a 7 French catheter (having an outer diameter of just over 2 mm) is provided with an ablation tip electrode at its distal end having a length ranging from about 4 mm to about 8 mm. If the ablation electrode is provided in perpendicular relation to the tissue, a relatively small surface area of the electrode is in contact with the tissue. In contrast, a relatively larger surface area would be in contact with the tissue if the ablation electrode were in a generally parallel relationship to the tissue, i.e., if the ablation electrode were positioned on its side. It is often difficult for the physician to determine the precise orientation of the tip electrode relative to the tissue because the tissue is not visible with fluoroscopy. The size of the lesion is often related to the amount surface area in contact with the tissue. As a result, there is less predictability in lesion size.

[0013] Thus, although it is desirable to have a relatively long electrode for cooling purposes, it is desirable to have a relatively small electrically active surface area for obtaining more discrete electrograms, for more accurately detecting char and thrombus formation, and for enhancing the predictability of lesion size, as well as to have a bipolar electrode arrangement that more accurately measures near-field activity.

SUMMARY OF THE INVENTION

[0014] The present invention is directed to a catheter that is particularly suitable for bipolar mapping and ablation. In one embodiment, the catheter comprises an elongated flexible body having a distal region and at least one lumen extending therethrough. A tip electrode is mounted on the distal region. The tip electrode has an exposed distal region having an outer diameter, a recessed central region having an outer surface proximal to the exposed distal region, and a proximal region having an outer diameter and an outer surface proximal to the central region. The recessed central region has an outer diameter less than the outer diameters of the exposed distal region and the proximal region. The central region and the proximal region are provided with an electrically insulating and thermally conductive layer over at least a portion of their outer surfaces. A ring electrode is mounted on the recessed central region. The ring electrode has an outer diameter less than the outer diameters of the exposed distal region and the proximal region.

[0015] With this design, the exposed region of the tip electrode is in direct contact with the heart tissue, and thus senses both the local activation energy (near-field signals) at the point of contact with the heart tissue and far field activation energy (far-field signals) received by the exposed region through the blood. However, the recessed electrode is protected from direct contact with the heart tissue, but does contact with surrounding blood. The close proximity of the recessed electrode to the exposed region enables the recessed electrode to receive approximately the same far-field signals as the exposed region. However, the recessed electrode does not pick up the local activation potential (near-field signals) that are received by the exposed region. Thus, this design permits the creation of high resolution electrograms.

[0016] In another embodiment, the catheter comprises an elongated flexible body having proximal and distal ends and at least one lumen extending therethrough. A tip section comprising a flexible tubing that is more flexible than the catheter body is mounted at the distal end of the catheter body. The flexible tubing has proximal and distal ends and at least on lumen extending therethrough. A tip electrode is mounted on the distal end of the flexible tubing. The tip electrode has an exposed distal region having an outer diameter, a recessed central region having an outer surface proximal to the exposed distal region, and a proximal region having an outer diameter and an outer surface proximal to the central region. The recessed central region has an outer diameter less than the outer diameters of the exposed distal region and the proximal region. The central region and the proximal region are provided with an electrically insulating and thermally conductive layer over at least a portion of their outer surfaces. A ring electrode is mounted on the recessed central region. The ring electrode has an outer diameter less than the outer diameters of the exposed distal region and the proximal region.

DESCRIPTION OF THE DRAWINGS

[0017] These and other features and advantages of the present invention will be better understood by reference to the following detailed description when considered in conjunction with the accompanying drawings wherein:

FIG. 1 is a side view of an embodiment of the catheter of the invention.

FIG. 2 is a side cross-sectional view of a catheter body according to the invention, including the junc-

tion between the catheter body and tip section.

FIG. 3A is a side cross-sectional view of a catheter tip section showing the tip electrode, ring electrode, and lumens for the sensor cable and puller wire.

FIG. 3B is a side cross-sectional view of FIG. 3A showing the tip electrode, ring electrode, and lumens for the sensor cable and thermocouple and electrode lead wires.

FIG. 3C is an enlarged view of the distal end of the thermocouple mounted in the tip electrode of the catheter of FIGs. 3A and 3B.

FIG. 4 is a longitudinal cross-sectional view of the tip section illustrated in FIGs. 3A and 3B.

FIG. 5 is a perspective view of the tip electrode depicted in FIGs. 3A and 3B.

FIG. 6 is a side view of the tip electrode depicted in FIGs. 3A and 3B.

DETAILED DESCRIPTION

[0018] In a particularly preferred embodiment of the invention, there is provided a steerable catheter having two electrodes for making bipolar measurements. As shown in FIGs. 1 and 2, catheter **10** comprises an elongated catheter body **12** having proximal and distal ends, a tip section **14** at the distal end of the catheter body **12**, and a control handle **16** at the proximal end of the catheter body **12**.

[0019] With reference to FIG. 2, the catheter body **12** comprises an elongated tubular construction having a single, axial or central lumen **18**. The catheter body **12** is flexible, i.e., bendable, but substantially non-compressible along its length. The catheter body **12** can be of any suitable construction and made of any suitable material. A presently preferred construction comprises an outer wall **22** made of a polyurethane, or PEBAX. The outer wall **22** comprises an imbedded braided mesh of high-strength steel, stainless steel or the like to increase torsional stiffness of the catheter body **12** so that, when the control handle **16** is rotated, the tip section **14** of the catheter **10** will rotate in a corresponding manner.

[0020] The outer diameter of the catheter body **12** is not critical, but is preferably no more than about 8 french, more preferably about 7 french. Likewise the thickness of the outer wall **22** is not critical, but is thin enough so that the central lumen **18** can accommodate a puller wire, lead wires, and any other wires, cables or tubes. The inner surface of the outer wall **22** is lined with a stiffening tube **20**, which can be made of any suitable material, such as polyimide or nylon. The stiffening tube **20**, along with the braided outer wall **22**, provides improved torsional stability while at the same time minimizing the wall thickness of the catheter, thus maximizing the diameter of the central lumen **18**. The outer diameter of the stiffening tube **20** is about the same as or slightly smaller than the inner diameter of the outer wall **22**. Polyimide tubing is presently preferred for the stiffening tube **20** because it may be very thin walled while still providing very good stiffness. This maximizes the diameter of the central lumen **18** without sacrificing strength and stiffness.

[0021] One preferred catheter has an outer wall **22** with an outer diameter of from about 0.090 inch to about 0.098 inch and an inner diameter of from about 0.061 inch to about 0.065 inch and a polyimide stiffening tube **20** having an outer diameter of from about 0.060 inch to about 0.064 inch and an inner diameter of from about 0.051 inch to about 0.056 inch. As would be recognized by one skilled in the art, the catheter body construction can be modified as desired. For example, the stiffening tube can be eliminated.

[0022] As shown in FIGs. 3A, 3B and 4, the tip section **14** comprises a short section of tubing **19** having three lumens **30**, **32** and **34**. The tubing **19** is made of a suitable non-toxic material that is preferably more flexible than the catheter body **12**. A presently preferred material for the tubing **19** is braided polyurethane, i.e., polyurethane with an embedded mesh of braided high-strength steel, stainless steel or the like. The outer diameter of the tip section **14**, like that of the catheter body **12**, is preferably no greater than about 8 french, more preferably about 7 french. The size of the lumens is not critical and can vary depending on the specific application. In a particularly preferred embodiment, the tip section **14** has an outer diameter of about 7 french (.092 inch) and the first lumen **30** and second lumen **32** are generally about the same size, each having a diameter of from about 0.020 inch to about 0.024 inch, preferably 0.022 inch, with the third lumen **34** having a slightly larger diameter of from about 0.032 inch to about 0.038 inch, preferably 0.036 inch.

[0023] A preferred means for attaching the catheter body **12** to the tip section **14** is illustrated in FIG. 2. The proximal end of the tip section **14** comprises an outer circumferential notch **24** that receives the inner surface of the outer wall **22** of the catheter body **12**. The tip section **14** and catheter body **12** are attached by adhesive (e.g., polyurethane glue) or the like. Before the tip section **14** and catheter body **12** are attached, however, the stiffening tube **20** is inserted into the catheter body **12**. The distal end of the stiffening tube **20** is fixedly attached near the distal end of the catheter body **12** by forming a glue joint (not shown) with polyurethane glue or the like. Preferably a small distance, e.g., about 3 mm, is provided between the distal end of the catheter body **12** and the distal end of the stiffening tube **20** to permit room for the catheter body **12** to receive the notch **24** of the tip section **14**. A force is applied to the proximal end of the stiffening tube **20**, and, while the stiffening tube **20** is under compression, a first glue joint (not shown) is made between the stiffening tube **20** and the outer wall **22** by a fast drying glue, e.g. Super Glue®. Thereafter a second glue joint (not shown) is formed between the proximal ends of the stiffening tube **20** and outer wall **22** using a slower drying but stronger glue, e.g., polyurethane.

[0024] At the distal end of the tip section **14** is a tip

electrode **36**. Preferably the tip electrode **36** generally has a diameter about the same as the outer diameter of the tubing **19**. The tip electrode **36** can be made from any suitable material, such as platinum, gold, or stainless steel, and is preferably made of a platinum-iridium alloy (90% platinum/10% iridium).

[0025] The tip electrode **36** is generally solid, having first and second blind holes **31** and **33** extending proximally therethrough, as described in more detail below. The tip electrode **36** has a length preferably ranging from about 4 mm to about 14 mm, more preferably from about 6 mm to about 12 mm. As depicted in FIGs. 3A, 3B, 5 and 6, the tip electrode **36** comprises an exposed distal region **35**, a proximal region **37**, and a recessed central region **38** between the exposed region **35** and the proximal region **37**. The exposed region **35** has a length preferably ranging from about 1.5 mm to about 4.0 mm, more preferably from about 2.0 mm to about 3.0 mm, still more preferably from about 2.3 mm to about 2.5 mm, and an outer diameter preferably ranging from about 2.0 mm to about 3.0 mm, more preferably from about 2.3 mm to about 2.5 mm.

[0026] In a particularly preferred arrangement, the exposed region has a length similar to its diameter. Preferably the ratio of the length to the diameter ranges from about 0.8:1 to about 1.2:1, more preferably from about 0.9:1 to about 1.1:1. In a particularly preferred embodiment, the length of the exposed region is approximately equal to the diameter of the exposed region. Thus, for example, for a 7 French electrode catheter, the tip electrode preferably has a length and diameter equal to about 2.3 mm. With such a design, the surface area of the tip electrode (having a generally hemispherical distal end) is the same in the parallel orientation and in the perpendicular orientation. These uniform dimensions provide a generally constant contact area between the exposed region **35** of the tip electrode **36** and the heart tissue, irrespective of the orientation of the tip section **14** relative to the heart tissue. As a result, more predictable lesion sizes can be achieved.

[0027] The central region **38** is recessed to permit a ring electrode **39** to be mounted thereon with the ring electrode being recessed relative to the exposed region **35** and proximal region **37** to be protected from direct contact with the heart tissue, but permitting contact with surrounding blood, as described in more detail below. Thus, the central portion **38** has an outer diameter that is smaller than the other diameter of the exposed region **35**, with the difference between the outer diameter of the central portion and the outer diameter of the exposed region preferably ranging from about 0.1 mm to about 1 mm, more preferably from about 0.5 mm to about 0.7 mm, thus resulting in a recess having a depth ranging from about 0.05 to about 0.4 mm, more preferably about 0.25 mm to about 0.35 mm. The central region **38** preferably has a length slightly greater than the length of the ring electrode **39** mounted thereon. The length of the central region **38** preferably ranges from about 0.5 mm to about 4 mm, more preferably from about 1 mm to about 3 mm.

[0028] The proximal region **37** has an outer diameter generally similar to the outer diameter of the exposed region **35**. The length of the proximal region **37** preferably ranges from about 3 mm to about 8 mm, more preferably from about 4 mm to about 6 mm.

[0029] The tip electrode **36** can be mounted on the distal end of the tubing **19** by any suitable technique. In the depicted embodiment, the proximal end of the proximal region **37** of the tip electrode **36** includes a circumferential notch **26** that mates with a corresponding circumferential notch **27** in the distal end of the tubing **19**. The tip electrode **36** is then bonded to the tubing **19** with polyurethane glue or the like. Alternatively, the proximal end of the tip electrode **36** can be provided with a stem, i.e., a region of reduced diameter, that fits into a hole in the distal end of the tubing **19**.

[0030] The proximal region **37** and central region **38** of the tip electrode **36** are provided with an electrically insulating and thermally conductive layer **55** over at least a portion, and preferably over all, of their outer surfaces. The electrically insulating and thermally conductive layer **55** is made of any suitable electrically insulating and thermally conductive material, including, but not limited to diamond, carbon, parylene, polyimide, polyester, polyurethane, epoxy, ceramic, and combinations thereof. The electrically insulating and thermally conductive layer **55** preferably has a thickness no greater than about 10 μm, more preferably ranging from about 0.5 μm to about 10 μm, still more preferably from about 1 μm to about 5 μm.

[0031] The insulating layer **55** can be applied by any suitable technique. For example, a layer of parylene can be applied by a deposition coating technique where the exposed region **35** of the tip electrode **36** is placed in a deposition chamber. The chamber functions to both hold the tip electrode **36** in place, as well as protect the exposed region **35** from exposure to parylene deposits. Alternatively, the insulating layer **55** may comprise a shrink-sleeve made of, for example, a thin polyester (e.g., about 0.00025 inch), or may comprise a thin coating of polyurethane or the like painted onto the surface of the central region **38**.

[0032] As described above, the ring electrode **39** is mounted on the central portion **38**. The ring electrode **39** is fixed in place by glue or the like over the portion of the electrically insulating and thermally conductive layer **55** on the recessed central portion **38**. The electrically insulating and thermally conductive layer **55** prevents electrical contact between the ring electrode **39** and the tip electrode **36**. The ring electrode **39** can be made of any suitable material, such as platinum, gold, or stainless steel, and is preferably made of a platinum-iridium alloy (90% platinum/10% iridium). The ring electrode **39** is preferably fixed near the distal end of the central portion **38** and in close proximity to the tip portion **35**.

[0033] The ring electrode can be created by any suitable technique. In one embodiment, the ring electrode **39** comprises a resilient ribbon-shaped conductive material that is wrapped within the recess **26** and fixed in place by glue or the like. The ring electrode **39** can be made of any suitable conductive material, such as those discussed above for the tip electrode. The width and thickness of the ring electrode **39** are suitable for fitting within the recess **26** so that the outer surface of the ring electrode **39** is recessed within the recess **26.** In other words, the ring electrode **39** has an outer diameter less than the outer diameter of the exposed region **35**. Preferably, the outer diameter of the ring electrode **39** is at least about 5%, more preferably from about 10% to about 20%, less than the outer diameter of the exposed region **35**, although the difference between the outer diameters could be greater if desired. The ring electrode **39** has a width preferably ranging from about 0.5 mm to about 4 mm, more preferably from about 1 mm to about 3 mm. In another embodiment, the ring electrode **39** is in the form of a snap ring, where the width and thickness of the ring **39** are suitable for fitting within the recess **26,** as described above.

[0034] The tip electrode **36** and ring electrode **39** are each connected to a separate lead wire **44**. The lead wires **44** extend through the first lumen **30** of tip section **14**, the central lumen **18** of the catheter body **12**, and the control handle **16**, and terminate at their proximal end in an input jack (not shown) that may be plugged into an appropriate signal processing unit and/or source of RF energy (not shown). The portion of the lead wires **44** extending through the central lumen **18** of the catheter body **12**, control handle **16** and proximal end of the tip section **14** may be enclosed within a protective sheath **49**, which can be made of any suitable material, preferably polyimide (as shown in FIGs. 2 and 4). The protective sheath **49** is preferably anchored at its distal end to the proximal end of the tip section **14** by gluing it in the first lumen **30** with polyurethane glue or the like.

[0035] The lead wires **44** are attached to the tip electrode **36** and ring electrodes **39** by any conventional technique. Connection of a lead wire **44** to the tip electrode **36** is accomplished, for example, by soldering the lead wire **44** into the second blind hole **33**.

[0036] The lead wire **44** for the ring electrode **39** extends through the first blind hole **31** in the tip electrode **36** and through an opening **25** in the side of the ring electrode, as shown in FIGs. 3A and 5. The ends of the lead wire **44** are then stripped of any coating and soldered or welded to the underside of the ring electrode **39**, which is positioned over the opening **25** and fixed in place with polyurethane glue or the like. Once the ring electrode **39** is fixed in place, extra polyurethane glue **28** is used to seal the region between the distal end of the proximal region **37** and the ring electrode in the general area of the opening **25** to protect internal components of the catheter from surrounding fluids.

[0037] Two additional ring electrodes **59** are mounted of the distal end the tip section **14**. The ring electrodes **59** are slid over the flexible tubing **19** and fixed in place by glue of the like. The position and spacing of the additional ring electrode **59** are not critical. If desired, more or less ring electrodes **59** may be used and can be positioned over the flexible tubing **19** of the tip section **14**.

[0038] A temperature sensing means is provided for the tip electrode **36** and, if desired, one or more of the ring electrodes **39** and **59**. Any conventional temperature sensing means, e.g., a thermocouple or thermistor, may be used. A preferred temperature sensing means for the tip electrode **36** comprises a thermocouple formed by an enameled wire pair. One wire of the wire pair is a copper wire **41**, e.g., a number **40** copper wire. The other wire of the wire pair is a constantan wire **45**. The wires **41** and **45** of the wire pair are electrically isolated from each other except at their distal ends where they are twisted together, covered with a short piece of plastic tubing **58**, e.g., polyamide, and covered with epoxy (as shown in FIG. 3c). The plastic tubing **58** is then attached in the distal end of the first blind hole **31** of the tip electrode **36** by polyurethane glue or the like. Alternatively, the wires **41** and **45** can be soldered in the blind hole **31**. The distal end of the first blind hole **31** preferably has a smaller diameter than the proximal end (which is sized to accommodate a location sensor, as described further below). The distal end of the blind hole **31** preferably has a size suitable for fitting the tubing **58** and wires **41** and **45** snugly into the blind hole. The distal end of the blind hole **31** is preferably positioned in the middle of the exposed region **35** to provide accurate temperature measurement by the thermocouple wires **41** and **45** mounted therein.

[0039] The wires **41** and **45** extend through the second lumen **32** in the tip section **14** and through the central lumen **18** of the catheter body **12**. The wires **41** and **45** then extend out through the control handle **16** and to a connector (not shown) connectable to a temperature monitor (not shown).

[0040] Alternatively, the temperature sensing means may be a thermistor. A suitable thermistor for use in the present invention is Model No. AB6N2-GC14KA143E/37C sold by Thermometrics (New Jersey).

[0041] Additionally, a location sensor **72**, preferably an electromagnetic location sensor, is contained within the distal end of the tip section **14**. The electromagnetic location sensor **72** is located within the proximal end of the first blind hole **31** in the tip electrode **36** and fixed by polyurethane glue or the like. The electromagnetic sensor **72** is connected to an electromagnetic sensor cable **74**, which extends through the third lumen **34** of the tip section **14**, through the catheter body **12**, and out through control handle **16**. The electromagnetic sensor cable **74** comprises multiple wires encased within a plastic covered sheath. The sensor cable **74** is connected to a circuit board, which amplifies the signal received from the electromagnetic sensor **72** and transmits it to a computer in a form understandable by the computer.

In a particularly preferred embodiment, the circuit board is contained within the control handle **16**, as described in U.S. Patent Application Serial No. 08/924,616, entitled "Steerable Direct Myocardial Revascularization Catheter", the disclosure of which is incorporated herein by reference. Suitable electromagnetic sensors for use with the present invention are described. for example, in EP-A-0 989 384
and U.S. Patent Nos. 5,558,091, 5,443,489, 5,480,422, 5,546,951, 5,568,809, and 5,391,199 and International Publication No. WO 95/02995, the disclosures of which are incorporated herein by reference. If desired, the sensor **72** can alternatively be contained, at least in part, within a rigid plastic housing, e.g., made of polyetheretherketone (PEEK), that is mounted between the tip electrode **36** and the flexible tubing **19**. Such a design is described in U.S. Patent No. 5,938,603, the disclosure of which is incorporated herein by reference.

[0042] Additionally, a mechanism is provided for deflecting the tip section **14**. In one embodiment, the mechanism comprises a puller wire **50** extending through the catheter body **12**. The puller wire **50** is anchored at its proximal end to the control handle **16** and is anchored at its distal end to the tip section **14**. The puller wire **50** is made of any suitable metal, such as stainless steel or Nitinol, and is preferably coated with Teflon® or the like. The coating imparts lubricity to the puller wire **50**. The puller wire **50** preferably has a diameter ranging from about 0.006 to about 0.010 inches.

[0043] A compression coil **52** is situated within the catheter body **12** in surrounding relation to the puller wire **50**. The compression coil **52** extends from the proximal end of the catheter body **12** to the proximal end of the tip section **14**. The compression coil **52** is made of any suitable metal, preferably stainless steel. The compression coil **52** is tightly wound on itself to provide flexibility, i.e., bending, but to resist compression. The inner diameter of the compression coil **52** is preferably slightly larger than the diameter of the puller wire **50**. The Teflon® coating on the puller wire **50** allows it to slide freely within the compression coil **52**. If desired, particularly if the lead wires **44** are not enclosed by a protective sheath **49**, the outer surface of the compression coil **52** can be covered by a flexible, non-conductive sheath **46**, e.g., made of polyimide tubing, to prevent contact between the compression coil **52** and any other wires within the catheter body **12**.

[0044] The compression coil **52** is anchored at its proximal end to the proximal end of the stiffening tube **20** in the catheter body **12** by glue joint **51** and at its distal end to the tip section **14** by glue joint **53**. Both glue joints **51** and **53** preferably comprise polyurethane glue or the like. The glue may be applied by means of a syringe or the like through a hole made between the outer surface of the catheter body **12** and the central lumen **18**. Such a hole may be formed, for example, by a needle or the like that punctures the outer wall **22** of the catheter body **12** and the stiffening tube **20** which is heated sufficiently to form a permanent hole. The glue is then introduced through the hole to the outer surface of the compression coil **52** and wicks around the outer circumference to form a glue joint about the entire circumference of the compression coil **52**.

[0045] The puller wire **50** extends into the second lumen **32** of the tip section **14**. The puller wire **50** is anchored at its distal end to the tip electrode **36** within the second blind hole **33** along with the lead wire **44**. A preferred method for anchoring the puller wire **50** within the tip electrode **36** is by crimping metal tubing **54** to the distal end of the puller wire **50** and soldering the metal tubing **54** inside the second blind hole **33**. Anchoring the puller wire **50** within the tip electrode **36** provides additional support for the tip electrode **36** on the flexible plastic tubing **19**, reducing the likelihood that the tip electrode **36** will separate from the tubing **19**. Alternatively, the puller wire **50** can be attached to the side of the tip section **14**. Such a design is described in U.S. Patent Application No. 08/924,611 (filed September 5, 1997), the disclosure of which is incorporated herein by reference. Within the second lumen **32** of the tip section **14**, the puller wire **50** extends through a plastic, preferably Teflon®, sheath **56**, which prevents the puller wire **50** from cutting into the wall of the tubing **19** when the tip section is deflected.

[0046] Longitudinal movement of the puller wire **50** relative to the catheter body **12**, which results in deflection of the tip section **14**, is accomplished by suitable manipulation of the control handle **16**. A suitable control handle design for use with the present invention is described in U.S. Patent Application No. 08/982,113, filed December 1, 1997, the disclosure of which is incorporated herein by reference. If desired, the catheter can be multidirectional, i.e., having two or more puller wires to enhance the ability to manipulate the tip section in more than one direction or to form two or more different curves. A description of such a design is provided in EP-A-0 904 796, EP-A-0 985 423, EP-A-0 982 407, EP-A-1 005 839 and EP-A-1 038 545, the disclosures of which is incorporated herein by reference.

[0047] The catheters in accordance with the present invention are particularly suitable for mapping the heart and ablating heart tissue, such as accessory signal pathways causing arrhythmias. In operation, the distal end of the catheter **10** is inserted into a vein or artery and advanced into the heart. To assist in positioning the tip section **14** of the catheter **10** at a desired position within the heart, the puller wire **50** and control handle **16** are used to deflect the tip section **14**. Once the tip section **14** has been positioned at or near the desired location of the heart tissue, the electrical activity of the heart may be identified, evaluated or mapped, and electrophysiological sources of arrhythmia may be identified and/or treated.

[0048] More specifically the bipolar electrode pair, namely, the exposed region **35** of the tip electrode **36** and ring electrode **39** mounted on the tip electrode, is

used to map the electrical activity of the heart. The catheter of the present invention is designed such that the exposed region **35** of the tip electrode **36** is in direct contact with the heart tissue. Thus, the exposed region **35** senses both the local activation energy (near-field signals) at the point of contact with the heart tissue and far field activation energy (far-field signals) received by the exposed region through the blood. However, the ring electrode **39** is recessed relative to the exposed region **35** to be protected from direct contact with the heart tissue, but permitting contact with surrounding blood. The close proximity of the ring electrode **39** to the exposed region **35** enables the ring electrode to receive approximately the same far-field signals as the exposed region. However, the ring electrode **39** does not pick up the local activation potential (near-field signals) that are received by the exposed region **35**. The signals received by the exposed region **35** and the ring electrode **39** are sent to a suitable signal processing unit. Within the signal processing unit, the signal detected by the ring electrode **39**, which includes only far-field activity, is subtracted from the signal detected by the exposed region **35** of the tip electrode **36**, which includes both near-field and far-field activity. Thus, the near-field signals can be more accurately determined. This improved method of detecting electrical activity allows the physician or operator to determine the location of the arrhythmiogenic focus more accurately for ablating and other purposes.

[0049] Additionally, the electrically insulating layer **55** limits the electrically active surface area of the tip electrode **36** to the exposed region **35**, providing improved electrogram resolution compared to a conventional tip electrode. More specifically, electrograms recorded from the relatively small exposed region **35** will be more discrete, reporting electrical activity from a smaller area of tissue.

[0050] The electromagnetic location sensor **72** in the tip section **14** of the catheter is used to provide a map of the heart so that the physician can better determine where to ablate and can have a better sense for where the tip section **14** is located during the procedure. In use, the patient is placed in a magnetic field generated, for example, by situating under the patient a pad containing coils for generating a magnetic field, and a reference electromagnetic sensor is fixed relative to the patient, e.g., taped to the patient's back. Signals generated in the heart by both the fixed reference sensor and the electromagnetic location sensor **72** in the catheter are amplified and transmitted to a computer that analyzes the signals and displays them on a monitor. By this method, the precise location of the location sensor **72** (and thus the tip section **14**, including the tip electrode **36**) relative to the reference sensor can be ascertained and visually displayed. The sensor **72** can also detect displacement of the catheter that is caused by contraction of the heart muscle.

[0051] The electromagnetic mapping sensor **72** preferably is used in combination with the tip electrode **36** and ring electrodes **39**. By combining the electromagnetic sensor **72** and electrodes **36** and **39**, a physician can simultaneously map the contours or shape of the heart chamber, the electrical activity of the heart, and the extent of displacement of the catheter. The tip electrode **36** and ring electrode **38** are used to monitor the strength of the electrical signals at a particular location. As would be recognized by one skilled in the art, the inventive tip electrode design can be effectively used without the inclusion of a location sensor, and thus the electromagnetic sensor **72** can be eliminated if desired.

[0052] Once the electrical activity of the heart is determined and mapped, if desired, the physician uses the same catheter for ablating the heart tissue. The relatively large surface area of the tip electrode **36**, which includes the exposed region **35** as well as the proximal region **37** and central region **38**, which are covered with the thermally conductive layer **55**, enhances the ability of the electrode to dissipate heat during ablation compared to a comparable tip electrode having a length closer to the that of the exposed region **35** of the present catheter. Compared to a standard 4 mm tip electrode, the tip electrode of the present invention having a preferred length of about 6 mm to about 8 mm provides improved convective cooling by the passing blood, such that additional power can be applied to the tissue. However, conventional 8 mm tip electrodes require more power than the inventive tip electrode. The tip electrode of the present invention preferably has an exposed length of only about 2 mm to about 3 mm, reducing the amount of power required to ablate compared to a tip electrode have an exposed length of 8 mm. Further, a conventional 8 mm tip electrode uses more energy heating the surrounding blood rather than heating the adjacent heart tissue. This heating of the blood can cause significant coagulation of the blood along the surface of the tip electrode. The present invention overcomes this problem by providing an electrically insulating, but thermally conductive layer over the non-exposed portion of the tip electrode **36**, allowing the electrical power to be concentrated at the exposed region **35**, while the insulated portion (the central region **38**) of the tip electrode is convectively cooled by passing blood.

[0053] As a result of this design, more power can be applied to the tissue with the inventive electrode compared to a standard tip electrode having a total length equal to the length of the exposed region **35** of the inventive electrode. This holds true because the blood can more efficiently cool the inventive electrode compared to the standard tip electrode. Additionally, the inventive electrode can create the same sized lesion as a standard tip electrode having a total length equal to the total length of the inventive electrode (but without an electrically insulating and thermally conductive covering), but using less power. This holds true because the inventive electrode has a higher current density than a standard tip electrode, while still having similar cooling capabilities, and therefore less power is needed.

[0054] Additionally, as discussed above, electrode impedance is routinely monitored during ablation to detect the potential formation of dangerous char or thrombus on the tip electrode. Because the electrically active area (i.e., the exposed region **35**) of the inventive tip electrode is smaller than with conventional ablation tip electrodes that offer similar cooling ability, changes in impedance due to char or thrombus formation are more readily detected with the inventive catheter.

[0055] The preceding description has been presented with reference to presently preferred embodiments of the invention. Workers skilled in the art and technology to which this invention pertains will appreciate that alterations and changes in the described structure may be practiced without meaningfully departing from the principal, spirit and scope of this invention.

[0056] Accordingly, the foregoing description should not be read as pertaining only to the precise structures described and illustrated in the accompanying drawings, but rather should be read consistent with and as support to the following claims which are to have their fullest and fair scope.

Claims

1. A catheter comprising:

   an elongated flexible body having a distal region and at least one lumen extending therethrough;
   a tip electrode mounted on the distal region, the tip electrode having an exposed distal region having an outer diameter, a recessed central region having an outer surface proximal to the exposed distal region, and a proximal region having an outer diameter and an outer surface proximal to the central region, wherein the recessed central region has an outer diameter less than the outer diameters of the exposed distal region and the proximal region, and further wherein the central region and the proximal region are provided with an electrically insulating and thermally conductive layer over at least a portion of their outer surfaces; and
   a ring electrode mounted on the recessed central region and having an outer diameter less than the outer diameters of the exposed distal region and the proximal region.

2. A catheter comprising:

   an elongated flexible body having proximal and distal ends and at least one lumen extending therethrough;
   a tip section comprising a flexible tubing that is more flexible than the catheter body mounted at the distal end of the catheter body, the flexible tubing having proximal and distal ends and at least one lumen extending therethrough;
   a tip electrode mounted on the distal end of the flexible tubing, the tip electrode having an exposed distal region having an outer diameter, a recessed central region having an outer surface proximal to the exposed distal region, and a proximal region having an outer diameter and an outer surface proximal to the central region, wherein the recessed central region has an outer diameter less than the outer diameters of the exposed distal region and the proximal region, and further wherein the central region and the proximal region are provided with an electrically insulating and thermally conductive layer over at least a portion of their outer surfaces; and
   a ring electrode mounted on the recessed central region and having an outer diameter less than the outer diameters of the exposed distal region and the proximal region.

3. A catheter as claimed in claim 1 or claim 2, wherein the tip electrode has a length ranging from 4 mm to 12 mm, preferably from 6 mm to 10 mm.

4. A catheter as claimed in any one of claims 1 to 3, wherein the exposed distal region has a length ranging from 1.5 mm to 4.0 mm, preferably from 2 mm to 3 mm.

5. A catheter as claimed in any one of claims 1 to 4, wherein the exposed region has a length approximately equal to its outer diameter.

6. A catheter as claimed in any one of claims 1 to 5, wherein the difference between the outer diameter of the central portion and the outer diameter of the exposed region ranges from 0.25 mm to 1.5 mm, preferably from 0.5 mm to 1 mm.

7. A catheter as claimed in any one of claims 1 to 6, wherein the electrically insulating and thermally conductive layer comprises a composition selected from the group consisting of diamond, carbon, parylene, polyimide, polyester, polyurethane, epoxy, ceramic, and combinations thereof.

8. A catheter as claimed in any one of claims 1 to 7, wherein the electrically insulating and thermally conductive layer has a thickness no greater than 10 μm, preferably from 1 μm to 5 μm.

9. A catheter as claimed in claim 2 or any claim dependent thereon, further comprising at least one ring electrode mounted on the distal end of the flexible tubing.

10. A catheter as claimed in claim 2 or any claim de-

pendent thereon, further comprising means for deflecting the tip section.



FIG. 1

FIG.2



FIG. 3A


35
36
55
39
38
25
31
55
33
37
54
72
44
27
26
34
32
50
59
74
14
59
56
19

FIG. 3B


3C
36
55
38
39
44
28
37
55
27
26
30
19
44
14
34
59
74
44
49

FIG. 3C



FIG. 4



FIG.5





FIG.6

European Patent Office

EUROPEAN SEARCH REPORT

Application Number
EP 01 30 5874

DOCUMENTS CONSIDERED TO BE RELEVANT

Category	Citation of document with indication, where appropriate, of relevant passages	Relevant to claim	CLASSIFICATION OF THE APPLICATION (Int.Cl.7)
A	WO 97 45157 A (JANSSEN W MICHAEL ;MANGANARO WILLIAM M (US); ADVANCED CORONARY INT) 4 December 1997 (1997-12-04) * page 31, line 26-33; figure 2 *	1	A61B18/14
A	US 5 447 519 A (PETERSON DAVID K) 5 September 1995 (1995-09-05) * column 6, line 64 - column 7, line 3; figure 1 *	1,2	
A	US 4 892 102 A (ASTRINSKY ELIEZER A) 9 January 1990 (1990-01-09) * abstract; figure 1 *	1,2	
A	US 5 487 757 A (JARACZEWSKI RICHARD S ET AL) 30 January 1996 (1996-01-30) * column 11, line 34-36; figure 7 *	1,2	
A	US 5 257 635 A (LANGBERG EDWIN) 2 November 1993 (1993-11-02) * column 6, line 30 *	1,2	TECHNICAL FIELDS SEARCHED (Int.Cl.7)
A	WO 93 15790 A (AVITALL BOAZ) 19 August 1993 (1993-08-19) * figure 4 *	1,2	A61B

The present search report has been drawn up for all claims

Place of search	Date of completion of the search	Examiner
THE HAGUE	19 September 2001	Papone, F

CATEGORY OF CITED DOCUMENTS

X : particularly relevant if taken alone
Y : particularly relevant if combined with another document of the same category
A : technological background
O : non-written disclosure
P : intermediate document

T : theory or principle underlying the invention
E : earlier patent document, but published on, or after the filing date
D : document cited in the application
L : document cited for other reasons

& : member of the same patent family, corresponding document

EPO FORM 1503 03.82 (P04C01)

ANNEX TO THE EUROPEAN SEARCH REPORT ON EUROPEAN PATENT APPLICATION NO. EP 01 30 5874

This annex lists the patent family members relating to the patent documents cited in the above-mentioned European search report. The members are as contained in the European Patent Office EDP file on
The European Patent Office is in no way liable for these particulars which are merely given for the purpose of information.

19-09-2001

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
WO 9745157 A	04-12-1997	US 5749914 A	12-05-1998
		AU 1077197 A	05-01-1998
		EP 0904126 A1	31-03-1999
		WO 9745157 A1	04-12-1997
US 5447519 A	05-09-1995	NONE	
US 4892102 A	09-01-1990	EP 0159753 A1	30-10-1985
US 5487757 A	30-01-1996	US 5545200 A	13-08-1996
US 5257635 A	02-11-1993	US 4945912 A	07-08-1990
		AT 150274 T	15-04-1997
		AU 4187893 A	23-09-1993
		AU 645486 B2	20-01-1994
		AU 4660189 A	26-06-1990
		CA 2003689 A1	25-05-1990
		DE 68927892 D1	24-04-1997
		EP 0445223 A1	11-09-1991
		EP 0745354 A2	04-12-1996
		WO 9006079 A1	14-06-1990
		US 5370644 A	06-12-1994
		US 5246438 A	21-09-1993
		US 5230349 A	27-07-1993
		CA 2070403 A1	04-02-1992
		EP 0497959 A1	12-08-1992
		WO 9202272 A1	20-02-1992
WO 9315790 A	19-08-1993	US 5354297 A	11-10-1994
		US 5327905 A	12-07-1994
		CA 2117487 A1	19-08-1993
		DE 69330120 D1	17-05-2001
		EP 0625919 A1	30-11-1994
		JP 7504834 T	01-06-1995
		WO 9315790 A1	19-08-1993
		US 6113556 A	05-09-2000
		US 5642736 A	01-07-1997
		CN 1082382 A	23-02-1994

EPO FORM P0459